



# Aflex Hose Ltd

World's Leading Manufacturer of PTFE Flexible Hose

## Certificate of Conformity

### USP Protocols

Aflex Hose Limited Natural PTFE hose products have been tested in accordance with USP protocols and are found to conform to the requirements of the following tests:

**USP 27, NF22, 2004, for Class VI Plastics at 70°C <88>**

- Systemic Injection Test
- Intracutaneous Test
- Implantation Test

**USP 27, NF22, 2004, Elution test <87> and are considered non-cytotoxic**

- L929 MEM Elution Test

The PTFE used in the manufacture of Aflex Hose Limited hose products is manufactured solely from materials which meet the requirements of USP Plastics Class VI.

Yours faithfully,

For and on behalf of Aflex Hose Limited

Richard C White  
Quality Assurance Manager

27<sup>th</sup> June 2005

The World's Leading Manufacturer of PTFE Flexible Hose

Spring Bank Industrial Estate, Watson Mill Lane.

Sowerby Bridge, Halifax HX6 3BW

Tel : (01422) 317200

Fax Nos : (01422) 836000 - UK Sales/Admin. (01422) 833000 - Export  
(01422) 836666 - Purchasing (01422) 839398 - Technical

Registered No. 1088141. Directors : B F Whitworth (Chairman), R S Anderson (Managing Director), A Herbert (Production Director)



CERTIFICATE No. 1977

Form 45